



# Certificate of Analysis

Catalog Number: TS-343  
Lot Number: 140902-246  
Toxin: 15 Acetyl DON  
Quantity: 5.0 ml  
Concentration or Concentration after reconstitution: 100.0µg/ml  
Reconstitution if applicable: To reconstitute use a volumetric pipette to accurately pipette 5.0 ml of Acetonitrile into each standard vial. Cap tightly and vortex (mix) for at least 30 seconds.  
Solvent: Acetonitrile  
Storage: 8°C or lower  
Date of Manufacturing: September 2, 2014  
Date of Expiration: 12 months before reconstitution from certificate date (see below) and 6 months after reconstitution

---

Purity: ≥ 98%  
Traceability: All components used in production of this standard can be tracked back to their source for each batch produced, allowing full product traceability.  
Testing Method: UV Spectroscopy  
Date of Expiration: HP 8452A Diode Array Spectrophotometer @221nm in Methanol  
Extinction Coefficient: 6836  
Molecular Weight: 338.4  
Standard Uncertainty (µg/ml): ±2.24  
Product Specification: 90-110 µg/ml  
Results (µg/ml): 97.1µg/ml

This is to certify that 15 Acetyl DON standard Lot number 140902-246 passes the required specifications and is released for sale.

Carrie Maune – Quality Manager

Date: September 3, 2014

Trilogy Analytical Laboratory, Inc.  
870 Vossbrink Drive Washington, MO 63090  
Phone: 636-239-1521 Fax: 636-239-1531 [www.trilogylab.com](http://www.trilogylab.com)

REVISION HISTORY

<b>REVISED DATE</b>	<b>REVISED BY</b>	<b>CHANGES MADE</b>
5/11/10	DANIELLE GANT	ORIGINAL DOCUMENT
6/24/10	Danielle Gant	Added page number and code
1/13/11	Danielle Grimes	Corrected Code Number
9/26/2011	Danielle Grimes	Added paginated numbers

Trilogy Analytical Laboratory, Inc.  
870 Vossbrink Drive Washington, MO 63090  
Phone: 636-239-1521 Fax: 636-239-1531 [www.trilogylab.com](http://www.trilogylab.com)