



# Certificate of Analysis

Catalog Number: TSL-317  
Lot Number: 180612- 302  
Toxin: Deoxynivalenol  
Quantity: 10.0 mL  
Concentration or  
Concentration after reconstitution: 100.0µg/mL  
Reconstitution if applicable: N/A  
Solvent: Methanol  
Storage: 8°C or lower  
Date of Manufacturing: June 12, 2018  
Date of Expiration: 1year from certificate date (see below).

---

Purity: ≥ 98%  
Traceability: All components used in production of this standard can be tracked back to their source for each batch produced, allowing full product traceability.  
Testing Method: UV Spectroscopy  
Testing Equipment Conditions: HP 8452A Diode Array Spectrophotometer  
Solvent: Methanol  
Wavelength: 219 nm  
E Value: 6208  
Standard Uncertainty (µg/mL): ±2.24  
Product Specification (µg/mL): 90-110  
Results (µg/mL): 98.3

This is to certify that DON standard Lot number 180612- 302 passes the required specifications and is released for sale.

Craig Humphrey – Chemistry Production Manager  
October 29, 2018