



# Certificate of Analysis

Catalog Number: TSL-343  
Lot Number: 180516-192  
Toxin: 15 Acetyl DON  
Quantity: 5.0 mL  
Concentration or Concentration after reconstitution: 100.0µg/mL  
Reconstitution if applicable: N/A  
Solvent: Acetonitrile  
Storage: 8°C or lower  
Date of Manufacturing: May 16, 2018  
Date of Expiration: 1 year from certificate date (see below)

---

Purity: ≥ 98%  
Traceability: All components used in production of this standard can be tracked back to their source for each batch produced, allowing full product traceability.  
Testing Method: UV Spectroscopy  
Testing Equipment Conditions: HP 8452A Diode Array Spectrophotometer @221nm in Methanol  
Extinction Coefficient: 6836  
Molecular Weight: 338.4  
Standard Uncertainty (µg/mL): ±2.24  
Product Specification (µg/mL): (90-110)  
Results (µg/mL): 98.8

This is to certify that 15 Acetyl DON standard Lot number 180516-192 passes the required specifications and is released for sale.

Craig Humphrey – Chemistry Production Manager  
July 11, 2018